

BOX ISOVISION -ABRALUX ASTM—ISO

The ISOVISION instrument is specially made to check the level of abrasion on glazes ceramic tiles. The inside of the instrument is painted in “neutral” as requested to the standards. To do the test, place the ISOVISION instrument in a dark room, and put inside at centre the abraded tile of 100x100mm surrounded by 8 tiles with the same dimension but no abraded. Write on the marked face of the r.p.m corresponding to the abrasion made. Observing the tile specimen from a distance of 2m and a height of 1.65 it is possible to see the difference between the tile abraded and the other tiles not abraded. To ensure that the test is true, the consensus of at least 3 observers is required.

TECHNICAL SPECIFICATIONS

- Fluorescent lamp 6000 ÷ 6500 °K - lighting of 300 lx
- Supply: 240V – 50/60 Hz single phase
- Internal Dimensions: 610x610x610mm
- Power: 18W
- Weight: 35kg



• **ACCESSORIES AND SPARE PARTS**

Fluorescent lamp—GT1540

Digital luxmeter range 0÷1.000lx—GT0116

Calibration Report by ACCREDIA for Digital Luxmeter—GT0117

